



Device Information

Device Code : RC-5+
Serial Number : EFI241106225
Mode Code : TLC30A08B

Probe Type : Temperature(internal)
Firmware Version : V5.4

Trip Information

Trip Id: 0000001
Description : Temperature recording.

Mark Event

N/A

Configuration Information

Start Mode : Immediately
Start Delay : 0s
Time Base : UTC +02:00
Repeat Start : Enable

Log Interval : 10s
Cycle Switch : Disable
Stop Mode : Manual + Software


Alarm Zone	Allow Time	Alarm Type	Total Time	Violations	Status
H1: Over 20.0 °C	0s	Sin	1m 40s	1	Alarm
Ideal Region	unlimited		2m 50s		
L1: Below 10.0 °C	0s	Sin	0s	0	OK

Logging Summary

Highest : 25.2 °C
Lowest : 11.7 °C
Average : 18.3 °C
MKT : 19.5 °C
First Alarm (Te): 2026/09/16 15:42:40

Start Time : 2026/09/16 15:42:40
Stop Time : 2026/09/16 15:47:10 (temporary)
Elapsed Time : 4m 30s
Data Points : 28

Normal — Over(Te) — Below(Te) — Error(Te)



Date	Time	°C	Date	Time	°C	Date	Time	°C	Date	Time	°C	Date	Time	°C
26/09/16	15:42:40	25.2												
26/09/16	15:42:50	25.2												
26/09/16	15:43:00	25.0												
26/09/16	15:43:10	24.9												
26/09/16	15:43:20	24.5												
26/09/16	15:43:30	24.0												
26/09/16	15:43:40	23.4												
26/09/16	15:43:50	22.8												
26/09/16	15:44:00	22.1												
26/09/16	15:44:10	21.4												
26/09/16	15:44:20	20.7												
26/09/16	15:44:30	20.0												
26/09/16	15:44:40	19.2												
26/09/16	15:44:50	18.5												
26/09/16	15:45:00	17.8												
26/09/16	15:45:10	17.2												
26/09/16	15:45:20	16.5												
26/09/16	15:45:30	15.8												
26/09/16	15:45:40	15.1												
26/09/16	15:45:50	14.5												
26/09/16	15:46:00	13.9												
26/09/16	15:46:10	13.2												
26/09/16	15:46:20	12.6												
26/09/16	15:46:30	12.0												
26/09/16	15:46:40	11.7												
26/09/16	15:46:50	11.7												
26/09/16	15:47:00	12.0												
26/09/16	15:47:10	12.4												